

Appln No. 10/622,221
Amdt date December 29, 2005
Reply to Office action of June 29, 2005

Amendments to the Claims:

This listing of claims will replace all prior versions, and listings, of claims in the application:

Listing of Claims:

Please amend claims 1, 2, 6, 12-14, 17 and 25-27 and cancel claims 5, 7, 8, 20 and 21.

1. (Currently Amended) A catheter for ablating tissue, the catheter comprising:
an elongated generally-tubular catheter body having proximal and distal ends; and
an electrode assembly at the distal end of the catheter body, the electrode assembly including a porous electrode arrangement that is generally transverse to the catheter body, the porous electrode arrangement comprising:

a non-conductive tubing mounted on the distal end of the catheter, a mid-section of the non-conductive tubing forming a curve that first bends away from and then back toward and past the axis of the catheter body forming a generally straight distal end of the non-conductive tubing;

one or more electrodes electrically ~~connected~~ connectable to a suitable energy source, the electrodes being mounted on the non-conductive tubing;

a porous sleeve mounted in surrounding relation to the one or more electrodes and defining an open space between the porous sleeve and the one or more electrodes; and

one or more irrigation openings fluidly connecting the open space to a lumen extending through the catheter through which fluid can pass;

wherein, in use, fluid passes through the lumen in the catheter, through the one or more irrigation openings, into the open space and through the porous sleeve.

2. (Currently Amended) The catheter according to claim 1, wherein the ~~electrode assembly further comprises a non-conductive tubing mounted on the distal end of the catheter over which the one or more electrodes are mounted, and wherein~~ non-conductive tubing includes

Appln No. 10/622,221
Amdt date December 29, 2005
Reply to Office action of June 29, 2005

at least one lumen fluidly connected to the lumen in the catheter body and to the one or more irrigation openings.

3. (Original) The catheter according to claim 2, wherein the one or more electrodes comprises a single coiled electrode wrapped around a portion of the non-conductive tubing.

4. (Original) The catheter according to claim 2, wherein the porous sleeve has proximal and distal ends that are bonded to the non-conductive tubing.

5. (Canceled)

6. (Currently Amended) The catheter according to claim 5, wherein the generally straight porous electrode arrangement forms an angle with the axis of the catheter body ranging from about 75° to about 110°.

7. (Canceled)

8. (Canceled)

9. (Original) The catheter according to claim 1, wherein the porous sleeve comprises expanded polytetrafluoroethylene.

10. (Original) The catheter according to claim 1, wherein the porous sleeve comprises expanded polytetrafluoroethylene that is expandable to no more than 10% at a distilled water flow rate of 30 to 40 cc/min.

Appln No. 10/622,221

Amdt date December 29, 2005

Reply to Office action of June 29, 2005

11. (Original) The catheter according to claim 1, wherein the porous sleeve comprises a material selected from the group consisting of porous nylon, sintered ceramics, woven meshes and cellular foam.

12. (Currently Amended) The catheter according to claim 1, wherein the porous electrode arrangement has a length ranging from about 10 to about 25 mm.

13. (Currently Amended) The catheter according to claim 1, wherein the porous electrode arrangement has a length ranging from about 10 to about 15 mm.

14. (Currently Amended) The catheter according to claim 1, wherein the electrode assembly further comprises one or more ring electrodes mounted proximal and/or distal to the porous electrode arrangement.

15. (Original) The catheter according to claim 1, wherein the electrode assembly further comprises one or more temperature sensors.

16. (Original) The catheter according to claim 15, wherein the one or more temperature sensors are mounted under the porous sleeve.

17. (Currently Amended) A catheter for ablating tissue, the catheter comprising:
an elongated generally-tubular catheter body having proximal and distal ends; and
an electrode assembly at the distal end of the catheter body, the electrode assembly comprising:

_____ a non-conductive tubing mounted on the distal end of the catheter body [[and]] having a lumen extending therethrough, the non-conductive tubing forming a curve that first bends away from and then back toward and past the axis of the catheter body forming a generally straight distal end of the non-conductive tubing, wherein the generally straight distal end of the non-

Appln No. 10/622,221
Amdt date December 29, 2005
Reply to Office action of June 29, 2005

conductive tubing forms an angle with the axis of the catheter body ranging from about 75° to about 110°; and

_____ a generally-straight porous electrode mounted on the generally straight distal end of the non-conductive tubing and being generally transverse to the axis of the catheter body, the porous electrode comprising:

one or more electrodes electrically ~~connected~~ connectable to a suitable energy source;

a porous sleeve mounted in surrounding relation to the one or more electrodes and defining an open space between the porous sleeve and the one or more electrodes; and

one or more irrigation openings fluidly connecting the open space to the lumen extending through the non-conductive tubing.

18. (Original) The catheter according to claim 17, wherein the one or more electrodes comprises a single coiled electrode wrapped around a portion of the non-conductive tubing.

19. (Original) The catheter according to claim 17, wherein the porous sleeve has proximal and distal ends that are bonded to the non-conductive tubing.

20. (Canceled)

21. (Canceled)

22. (Original) The catheter according to claim 17, wherein the porous sleeve comprises expanded polytetrafluoroethylene.

23. (Original) The catheter according to claim 17, wherein the porous sleeve comprises expanded polytetrafluoroethylene that is expandable to no more than 10% at a distilled water flow rate of 30 to 40 cc/min.

Appln No. 10/622,221
Amdt date December 29, 2005
Reply to Office action of June 29, 2005

24. (Original) The catheter according to claim 17, wherein the porous sleeve comprises a material selected from the group consisting of porous nylon, sintered ceramics, woven meshes and cellular foam.

25. (Currently Amended) The catheter according to claim 17, wherein the porous electrode arrangement has a length ranging from about 10 to about 25 mm.

26. (Currently Amended) The catheter according to claim 17, wherein the porous electrode arrangement has a length ranging from about 10 to about 15 mm.

27. (Currently Amended) The catheter according to claim 17, wherein the electrode assembly further comprises one or more ring electrodes mounted proximal and/or distal to the porous electrode arrangement.

28. (Original) The catheter according to claim 17, wherein the electrode assembly further comprises one or more temperature sensors.

29. (Original) The catheter according to claim 28, wherein the one or more temperature sensors are mounted under the porous sleeve.

30. (Original) The catheter according to claim 28, further comprising a pre-shaped support wire extending through a second lumen in the non-conductive tubing.

31. (Original) The catheter according to claim 30, wherein the pre-shaped support wire comprises nitinol.

Appln No. 10/622,221
Amdt date December 29, 2005
Reply to Office action of June 29, 2005

32. (Original) The catheter according to claim 17, wherein the one or more irrigation openings are located only on the side of the porous electrode that is to be in contact with tissue to be ablated.

33. (Withdrawn) A method for ablating heart tissue comprising inserting the distal end of a catheter according to claim 1 into the heart of a patient and forming at least one linear lesion in the atrial tissue with the porous electrode by simultaneously supplying ablation energy to the one or more electrodes and introducing fluid through the irrigation openings so that the fluid passes through the porous sleeve.

34. (Withdrawn) A method for treating atrial fibrillation comprising inserting the distal end of a catheter according to claim 1 into an atrium of the heart of a patient and forming at least one linear lesion in the atrial tissue with the porous electrode by simultaneously supplying ablation energy to the one or more electrodes and introducing fluid through the irrigation openings so that the fluid passes through the porous sleeve.